


Betriebsdaten:
 Typical operation:
 Exemple d'utilisation:

U_f . . . 4 V~
 I_f . . . 2,3 A
 U_a . max 7000 Veff.
 I_a . . max 5 mA



Prüfdaten:
 Test operation:
 Vérification:

U_f . . . 4 V~
 U_a . . 100 V~

393

Patent-Röhrenprüfer Modell W19

MAX FUNKE, Adenau-Eifel 9, 54.

Valvo
 1875
 1875 P-01



Philips
 1875
 1775 P-01

Einweg-Gleichrichter / Half wave rectifier / Redresseur monoplaque.

Unbrauchbar
 mauvais / Bad

?

G u t
 Good / Bon



10 mA