


12 A 8 Heptode

12 A 8 G

12 A 8 GT



Hauptinstrument

Bereich

● **6mA**

● A 250 Volt

● U_{II} 3 Volt

Sollwert: 3,5mA

G_4 ●

● U_{IV} 100 Volt

● G_{3+5}

●

●

●

● U_{II} 200 Volt

● G_2

● U_I 7 Volt

● G_1

● M

● K

● F

Heizung 12,6 Volt

● F

Prüfkarte für RPM 370

Neuberger, München 25