



# 6 BW 7 Pentode



$J_A$  - Bereich

A

30

180 V

$J_A$  9,5 mA

$G_2$

## Meßspannungen

|           |       |     |    |
|-----------|-------|-----|----|
| Heizung   | F     | 6,3 | V  |
| $U_I$     | $G_1$ | 1,4 | -V |
| $U_{II}$  |       |     | -V |
| $U_{III}$ | $G_2$ | 180 | V  |
| $U_{IV}$  |       |     | V  |
| A         |       | 180 | V  |

## Richtströme

|           |     |      |
|-----------|-----|------|
| $J_A$     | 9,5 | mA   |
| $J_{G_1}$ | 3,5 | mA   |
| S         | 9,3 | mA/V |

$U_I$

$U_{III}$

180 V

$G_1$

K

K

F

F

$G_3$

Heizung 6,3 Volt

Meßkarte für NEUBERGER Röhrenmeßgeräte