


UF 21 Pentode



Hauptinstrument
Bereich

30mA ●

● A 200 Volt

Sollwert: 6mA

● ● ●

● U_{III} 100 Volt

● G₂

● U_I 2,5 Volt

● G₁

● G₃

● K

F ●

Heizung: 12,6 Volt / 100 mA

● F

Prüfkarte für RPM 370
Neuberger, München 25