


VF 7

Pentode



Hauptinstrument
Bereich

● **6mA**

● *A 200 Volt*

Sollwert : 3mA

● *U_{III} 100 Volt*

● *G₂*

● *U_I 2 Volt*

G₁ ●

G₃ ●

K ●

F ●

Heizung : 55 Volt / 50 mA

● *F*

Prüfkarte für RPM 370
Neuberger, München 25